

Alarm ✘

ID: TZ0724080307

File Information

App(Windows):v1.3.41

File Created Date: 04/29/2025 22:57:40

Note: All times shown are based on UTC+01:00 and 24-Hour clock [MM/DD/YYYY HH:MM:SS]

Device Information

Device Type:	TZ-TempU07B	Firmware Version:	1.04
ID:	TZ0724080307		
Start Delay:	30 M	Log Interval:	600 S
Description:	Temperature and Humidity Recorder		

Logging Summary

First Point:	08/27/2024 10:48:53	Max:	24.3°C(Temp)/71.5%RH(Humidity)
Stop Time:	10/12/2024 10:28:53	Min:	18.6°C(Temp)/57.2%RH(Humidity)
Number of Points:	6623	Average:	21.3°C(Temp)/64.2%RH(Humidity)
Trip Length:	45d 23h 40m 0s	MKT:	21.3°C
Stop Mode:	USB Stop	Start Mode:	Button Start

Marked Events

N/A

Alarm Condition	Alarm Delay	Alarm Type	First Point Time	Time of Violations	No. of Violations	Status
H1:Over 8.0°C	00:10:00	Single	08/27/2024 10:58:53	45.23:40:00	1	Alarm
L1:Below 2.0°C	00:10:00	Single	00/00/00 00:00:00	00:00:00	0	OK
H1:Over 80.0%	00:10:00	Single	00/00/00 00:00:00	00:00:00	0	OK
L1:Below 40.0%	00:10:00	Single	00/00/00 00:00:00	00:00:00	0	OK





























